

Sample Name: CEF3
 Cat No. : RP11765
 Lot No. : E1576621G
 Time Processed : 3:39:13
 Month-Day-Year Processed : 07/13/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump B Conc.	5
25.00	Pumps	Pump B Conc.	65
25.01	Pumps	Pump B Conc.	95
27.00	Pumps	Pump B Conc.	95
27.01	Pumps	Pump B Conc.	5
32.00	Pumps	Pump B Conc.	5
32.01	Controller	Stop	

<<Column Performance>>


<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Equipment: GK12010012

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	11.349	32947	3718	0.575
2	11.591	10192	1085	0.178
3	12.244	5664053	544801	98.873
4	15.439	3656	410	0.064
5	20.452	17742	801	0.310
Total		5728590	550815	100.000