

Sample Name : OVA Peptide (257-264)  
 Cat No. : RP10611  
 Lot No. : PE0821711

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow: 1 ml/min  
 Wavelength: 220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump B Conc.	5
25.00	Pumps	Pump B Conc.	65
25.01	Pumps	Pump B Conc.	95
31.00	Pumps	Pump B Conc.	95
31.01	Pumps	Pump B Conc.	5
40.00	Pumps	Pump B Conc.	5
40.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Alltima™ C18 4.6 x 250 mm

### <Chromatogram>

mV



### <Peak Table>

Detector A 220nm

Peak#	Ret. Time	Area	Height	Area%
1	13.097	1502	257	0.031
2	13.408	5599	959	0.115
3	13.571	80669	9231	1.660
4	13.770	42265	6454	0.870
5	13.981	4682990	769268	96.380
6	14.469	42558	7341	0.876
7	14.727	3299	554	0.068
Total		4858884	794065	100.000