

Sample Name :N-Formyl-Met-Leu-Phe
 Cat. No. :RP12956
 Lot. No. :PE2731806

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump B Conc.	5	
25.00	Pumps	Pump B Conc.	65	
25.01	Pumps	Pump B Conc.	95	
27.00	Pumps	Pump B Conc.	95	
27.01	Pumps	Pump B Conc.	5	
35.00	Pumps	Pump B Conc.	5	
35.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	17.133	99882	11309	0.703
2	18.595	40830	7384	0.288
3	19.526	5964	888	0.042
4	19.682	9139	1570	0.064
5	19.957	145646	23910	1.026
6	20.292	86765	9652	0.611
7	20.780	13754288	1593443	96.868
8	21.095	23086	4239	0.163
9	21.328	1214	174	0.009
10	21.674	32243	4368	0.227
Total		14199057	1656937	100.000