

Sample Name :Bombesin
 Cat. No. : RP11157
 Lot. No. : E9608819H
 Time Processed : 2:17:55
 Month-Day-Year Processed :09/06/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
33.00	Pumps	Solvent B Conc.	5
33.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment:ZJ17010507

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	10.940	1800	355	0.078
2	12.883	2675	569	0.116
3	13.175	1734	233	0.075
4	13.408	4835	779	0.210
5	14.002	9563	1650	0.415
6	15.037	2899	464	0.126
7	15.350	4143	565	0.180
8	15.595	2267148	390894	98.388
9	16.231	3386	653	0.147
10	17.310	2445	206	0.106
11	17.834	2565	225	0.111
12	24.022	1109	197	0.048
Total		2304302	396790	100.000