

Sample Name : LCMV gp33-41
 Cat No.: RP20257
 Lot No.: E90372191
 Time Processed : 18:57:55
 Month-Day-Year Processed : 09/06/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm

<<LC Time Program>>


Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	20
25.00	Pumps	Solvent B Conc.	80
25.01	Pumps	Solvent B Conc.	95
31.00	Pumps	Solvent B Conc.	95
31.01	Pumps	Solvent B Conc.	20
40.00	Pumps	Solvent B Conc.	20
40.01	Controller	Stop	

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ19010324

<Chromatogram>



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	9.904	63057	8351	0.299
2	14.354	743099	89867	3.526
3	14.708	20050969	1902852	95.152
4	14.983	132859	23351	0.630
5	16.085	82545	10234	0.392
Total		21072529	2034654	100.000