

Sample Information

Sample Name :HEL(46-61)  
 Cat. No. : RP11736  
 Lot. No. : PE5761805

Pump A : 0.065% trifluoroacetic in 100% waters (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)

<<Pump>>

Flow : 1.0000 mL/min

<<Detector A>>


Wavelength Ch1 : 220 nm

<<LC Time Program>>

Time	Module	Command	Value	Comment
0.01	Pumps	Solvent B Conc.	5	
25.00	Pumps	Solvent B Conc.	65	
25.01	Pumps	Solvent B Conc.	95	
27.00	Pumps	Solvent B Conc.	95	
27.01	Pumps	Solvent B Conc.	5	
35.00	Pumps	Solvent B Conc.	5	
35.01	Controller	Stop		

<<Column Performance>>

<Detector A>  
 Column : Inertsil ODS-3 4.6 x 250 mm



Peak Table

Peak#	Ret. Time	Height	Area	Height%	Area%
1	11.704	1097	4183	0.288	0.156
2	11.873	2166	12292	0.569	0.459
3	12.322	345332	2551045	90.630	95.240
4	12.504	32441	111022	8.514	4.145
Total		381037	2678541	100.000	100.000