

Sample Name :LL-37, scrambled  
 Cat. No. : RP20332  
 Lot. No. : E7107421A  
 Time Processed :5:39:56  
 Month-Day-Year Processed :01/15/2021  
 Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm  
 Equipment: SS-CM-0309

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	9.345	1303	173	0.029
2	12.782	2087	319	0.047
3	13.683	1538	135	0.035
4	14.148	5209	370	0.118
5	14.262	2501	370	0.057
6	14.372	2317	378	0.052
7	14.513	4092	447	0.093
8	14.704	3857	465	0.087
9	15.502	4310593	493133	97.501
10	15.735	42296	8773	0.957
11	16.407	2982	343	0.067
12	16.604	4800	725	0.109
13	16.922	2052	320	0.046
14	17.158	9724	1224	0.220
15	18.186	2540	330	0.057
16	19.312	4565	486	0.103

Peak#	Ret. Time	Area	Height	Area%
17	20.751	4580	331	0.104
18	22.007	1872	240	0.042
19	23.498	10318	619	0.233
20	24.611	1842	219	0.042
Total		4421068	509400	100.000