

Sample Name :Pep-1 (uncapped)
 Cat No: RP10130
 Lot No : E2191121C
 Time Processed :0:37:19
 Month-Day-Year Processed :03/10/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)

Total Flow:1 ml/min

Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
31.00	Pumps	B.Conc	95
31.01	Pumps	B.Conc	5
40.00	Pumps	B.Conc	5
45.00	Controller	Stop	

<<Column Performance>>


<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm

Equipment: ZJ19010140

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	10.036	4903880	669214	95.019
2	10.273	67373	10370	1.305
3	10.363	10512	2614	0.204
4	10.584	107113	15909	2.075
5	10.894	41873	5685	0.811
6	10.990	24550	6529	0.476
7	14.970	5646	1031	0.109
Total		5160947	711352	100.000