


Mass Spectrum



Sample Information

Month-Day Processed : 03/20/21
Time Processed : 4:33:36 PM
Injection Volume : 0.5
Sample Name : Luteinizing Hormone Releasing Hormone (LH-RH), salmon
Cat No. : RP10673
Lot No. : E9934021C
Theoretical MW : 1212.31
Observed MW : 1211.5

Interface : ESI
Nebulizing Gas Flow : 1.5L/min
CDL Temp : 250
Block Temp : 200

Equipment : GK11010007
Interface Bias : +3.5 kV
Drying Gas Flow : 5 L/min
T.Flow : 0.2 ml/min
B.conc : 50%H2O/50%MeOH