

Sample Name :GP (33-41)
 Cat No. :RP20091
 Lot No. :E7166820L
 Time Processed :8:41:02 AM
 Month-Day-Year Processed :12/26/20

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ17010508

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	8.672	5429	435	0.076
2	9.910	1603	305	0.022
3	12.238	26132	4287	0.364
4	12.518	5656	869	0.079
5	12.669	35649	6119	0.497
6	12.965	32106	3805	0.448
7	13.116	110404	11494	1.539
8	13.369	8495	1965	0.118
9	13.667	6858833	1032428	95.602
10	13.975	3249	1221	0.045
11	14.538	7360	1284	0.103
12	14.866	67621	9036	0.943
13	15.655	1035	181	0.014
14	16.773	6889	1195	0.096
15	16.965	1711	308	0.024

Peak#	Ret. Time	Area	Height	Area%
16	17.322	2160	236	0.030
Total		7174332	1075168	100.000