

Sample Name : LCMV gp33-41

Cat No.: RP20257

Lot No.: PE6211709

Pump A : 0.065% trifluoroacetic in 100% water (v/v)

Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)

Total Flow: 1 ml/min

Wavelength: 220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
31.00	Pumps	Pump A B.Conc	95	
31.01	Pumps	Pump A B.Conc	5	
40.00	Pumps	Pump A B.Conc	5	
40.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	11.672	6335	1149	0.205
2	13.254	4870	892	0.158
3	13.423	44002	6260	1.426
4	13.675	7475	1019	0.242
5	15.345	17105	2264	0.554
6	15.595	2934015	433704	95.079
7	15.971	2409	464	0.078
8	16.298	21488	2716	0.696
9	16.478	5090	662	0.165
10	17.183	14184	1862	0.460
11	17.267	5167	1353	0.167
12	19.009	14166	2126	0.459
13	20.175	9572	1290	0.310
Total		3085879	455758	100.000