

Sample Name :Bombesin
 Cat No.: RP11157
 Lot No.: E8555720F
 Time Processed :15:16:52
 Month-Day-Year Processed :06/05/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	15
25.00	Pumps	Solvent B Conc.	75
25.01	Pumps	Solvent B Conc.	95
31.00	Pumps	Solvent B Conc.	95
31.01	Pumps	Solvent B Conc.	15
40.00	Pumps	Solvent B Conc.	15
40.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment:ZJ19010324

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	10.196	25757	3743	0.247
2	10.396	41851	4629	0.401
3	10.619	200613	39270	1.921
4	10.773	10008799	1600132	95.830
5	10.900	79941	39334	0.765
6	13.757	87408	11306	0.837
Total		10444368	1698413	100.000