

Sample Name : [Arg8]-Vasopressin (AVP)  
 Cat No.: RP20275  
 Lot No.: E2739220K  
 Time Processed : 11:27:55  
 Month-Day-Year Processed : 11/29/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow: 1 ml/min  
 Wavelength: 220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	0
5.00	Pumps	Solvent B Conc.	0
20.00	Pumps	Solvent B Conc.	30
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
31.00	Pumps	Solvent B Conc.	95
31.01	Pumps	Solvent B Conc.	0
40.00	Pumps	Solvent B Conc.	0
40.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm  
 Equipment: ZJ19010014

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	19.931	23706	3261	0.257
2	20.558	78	39	0.001
3	20.970	1655	302	0.018
4	21.094	3398	516	0.037
5	21.391	8753982	1032306	95.081
6	21.680	302366	39549	3.284
7	21.930	121683	14699	1.322
Total		9206867	1090672	100.000