

Sample Name :N-Boc-Phe-Leu-Phe-Leu-Phe  
 Cat No.: RP12950  
 Lot No.: E0497919L  
 Time Processed :8:36:07  
 Month-Day-Year Processed :12/29/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	5
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
27.00	Pumps	Pump A B.Conc	95
27.01	Pumps	Pump A B.Conc	5
35.00	Pumps	Pump A B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm  
 Equipment:ZJ19010324

### <Chromatogram>

mV



### <Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	5.594	16461	2806	0.698
2	15.470	1754	191	0.074
3	16.409	3955	239	0.168
4	17.081	3109	473	0.132
5	17.968	3867	508	0.164
6	18.079	1104	215	0.047
7	18.257	2220	320	0.094
8	18.359	2017	363	0.086
9	18.600	17121	4297	0.726
10	18.750	2277216	351818	96.603
11	19.421	4343	575	0.184
12	20.269	1362	298	0.058
13	20.702	1915	198	0.081
14	20.913	3030	465	0.129
15	20.984	2272	508	0.096

Peak#	Ret. Time	Area	Height	Area%
16	21.084	2624	378	0.111
17	21.899	1393	181	0.059
18	22.801	1596	290	0.068
19	23.016	1519	245	0.064
20	23.354	1650	276	0.070
21	23.930	1913	361	0.081
22	24.165	1154	283	0.049
23	24.500	2416	237	0.102
24	24.808	1285	245	0.055
Total		2357297	365767	100.000