

Sample Name : C3a (70-77)
 Cat. No. : RP20511
 Lot. No. : E2686520K
 Time Processed : 10:38:15 PM
 Month-Day-Year Processed : 12/04/2020
 Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	0
25.00	Pumps	Pump A B.Conc	60
25.01	Pumps	Pump A B.Conc	95
31.00	Pumps	Pump A B.Conc	95
31.01	Pumps	Pump A B.Conc	0
40.00	Pumps	Pump A B.Conc	0
40.00	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment : ZJ19010140

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	13.454	123899	8580	1.140
2	13.697	10404830	1140044	95.776
3	13.871	107413	26369	0.989
4	13.994	88629	16496	0.816
5	14.079	44684	5008	0.411
6	18.908	94205	10923	0.867
Total		10863658	1207420	100.000