

Sample Name :Astressin
 Cat No :RP10980
 Lot No :E4293720L
 Time Processed :10:03:04 AM
 Month-Day-Year Processed :12/27/2020


Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

Time	Module	Command	Value
0.10	Pumps	Pump A B.Conc	5
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
27.00	Pumps	Pump A B.Conc	95
27.01	Pumps	Pump A B.Conc	5
70.00	Pumps	Pump A B.Conc	5
72.00	Controller	Stop	

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm



Peak Table

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	14.744	41830	5855	1.070
2	15.854	5462	1103	0.140
3	16.009	3797634	574743	97.137
4	16.221	17805	4282	0.455
5	16.497	22361	2457	0.572
6	17.513	24453	3803	0.625
Total		3909545	592243	100.000