

Sample Name :GP (33-41)
 Cat No. : RP20091
 Lot No. : E0319920B
 Time Processed : 20:46:32
 Month-Day-Year Processed : 02/10/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	5
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
31.00	Pumps	Pump A B.Conc	95
31.01	Pumps	Pump A B.Conc	5
40.00	Pumps	Pump A B.Conc	5
40.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment:GK12010012

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	13.543	11485	1831	0.214
2	13.992	45529	8632	0.846
3	14.145	5203560	806401	96.744
4	14.292	95511	24147	1.776
5	14.865	21681	2349	0.403
6	24.636	940	125	0.017
Total		5378706	843485	100.000