

Sample Name :Substance P  
 Cat No. :RP10178  
 Lot No. :E1306620E  
 Time Processed :8:44:57  
 Month-Day-Year Processed :05/07/20

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
35.00	Pumps	Solvent B Conc.	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm  
 Equipment: ZJ17010508

### <Chromatogram>

mV



### <Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	14.410	45940	3466	0.680
2	15.855	6546426	957299	96.872
3	16.109	65421	10279	0.968
4	16.409	33300	5081	0.493
5	18.083	46347	5188	0.686
6	19.182	2337	392	0.035
7	20.525	4558	514	0.067
8	21.631	13504	1429	0.200
Total		6757834	983648	100.000