

Sample Name :GP (33-41)  
 Cat. No. : RP20091  
 Lot. No. : E5078220J  
 Time Processed :3:22:39 AM  
 Month-Day-Year Processed :10/14/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
31.00	Pumps	B.Conc	95
31.01	Pumps	B.Conc	5
40.00	Pumps	B.Conc	5
40.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm  
 Equipment: ZJ19010140

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	8.467	8813	1665	0.057
2	10.721	47289	3305	0.304
3	10.922	15466230	2016145	99.316
4	11.165	50480	6328	0.324
Total		15572811	2027444	100.000