

Sample Name : β -Amyloid (1-42), rat
 Cat. No. : RP10013
 Lot. No. : PE1661704

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow : 1 ml / min
 Wavelength : 220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump B Conc.	10	
25.00	Pumps	Pump B Conc.	70	
25.01	Pumps	Pump B Conc.	95	
31.00	Pumps	Pump B Conc.	95	
31.01	Pumps	Pump B Conc.	10	
40.00	Pumps	Pump B Conc.	10	
40.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : AlltimaTM C18 4.6 x 250 mm

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	15.654	14542	2361	0.254
2	15.932	12426	1955	0.217
3	16.503	5456669	363005	95.170
4	17.215	130083	16246	2.269
5	17.360	26439	7055	0.461
6	17.803	64290	9530	1.121
7	18.151	4375	1271	0.076
8	18.605	10728	1924	0.187
9	19.515	14033	2363	0.245
Total		5733585	405710	100.000