

Sample Name : OVA Peptide (323-339)
 Cat. No. :RP10610
 Lot. No. :E7230620B
 Time Processed : 9:36:49
 Month-Day-Year Processed : 02/09/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump B Conc.	5	
25.00	Pumps	Pump B Conc.	65	
25.01	Pumps	Pump B Conc.	95	
27.00	Pumps	Pump B Conc.	95	
27.01	Pumps	Pump B Conc.	5	
35.00	Pumps	Pump B Conc.	5	
35.01	Controller	Stop		


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm

Equipment: SS-CM-0309

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	10.042	7655	1564	0.460
2	10.336	10113	1715	0.608
3	10.498	1588139	323054	95.424
4	11.181	9549	1723	0.574
5	14.878	26881	2892	1.615
6	16.778	21963	1286	1.320
Total		1664301	332234	100.000