

Sample Name :GRGDSPC
 Cat No. : RP20283
 Lot No. :PE1921903
 Time Processed : 19:37:29
 Month-Day-Year Processed : 03/15/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump B Conc.	5
25.00	Pumps	Pump B Conc.	65
25.01	Pumps	Pump B Conc.	95
27.00	Pumps	Pump B Conc.	95
27.01	Pumps	Pump B Conc.	5
32.00	Pumps	Pump B Conc.	5
32.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment:GK12010012

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	6.254	4124	649	0.044
2	6.475	6014	949	0.064
3	6.831	8997718	1270167	95.875
4	7.067	212018	32565	2.259
5	7.220	92334	10099	0.984
6	7.714	38126	3645	0.406
7	8.023	10443	1549	0.111
8	9.070	4551	573	0.048
9	11.035	1118	87	0.012
10	13.001	1159	230	0.012
11	13.384	1553	216	0.017
12	14.410	1839	193	0.020
13	15.436	1412	172	0.015
14	16.214	9024	1098	0.096
15	22.175	2152	206	0.023

Peak#	Ret. Time	Area	Height	Area%
16	24.077	1222	265	0.013
Total		9384807	1322662	100.000