

Sample Name : GP (33-41)  
 Cat. No. : RP20091  
 Lot. No. : PE2871809  
 Time Processed :19:43:06  
 Month-Day-Year Processed :09/30/2018

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
27.00	Pumps	Pump A B.Conc	95	
27.01	Pumps	Pump A B.Conc	5	
35.00	Pumps	Pump A B.Conc	5	
35.01	Controller	Stop		


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Equipment: GR11010440

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	12.915	11813	1994	0.290
2	13.139	7147	1031	0.176
3	13.350	2013	294	0.049
4	13.634	3749	654	0.092
5	14.090	18515	2652	0.455
6	14.324	3914567	571245	96.136
7	14.569	7324	1855	0.180
8	14.806	96769	17260	2.377
9	15.111	5949	819	0.146
10	16.011	2492	364	0.061
11	20.337	1553	261	0.038
Total		4071890	598429	100.000