

Sample Name : OVApeptide(257-264)  
 Cat. No. : RP10611  
 Lot. No. : PE3351704

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
31.00	Pumps	Pump A B.Conc	95	
31.01	Pumps	Pump A B.Conc	5	
40.00	Pumps	Pump A B.Conc	5	
40.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Alltima™ C18 4.6 x 250 mm

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	12.522	4450	882	0.090
2	13.356	3102	654	0.062
3	14.448	14138	2610	0.285
4	14.547	5829	1202	0.117
5	14.888	46200	8532	0.931
6	16.220	67100	6253	1.352
7	16.492	50996	8037	1.027
8	16.629	4729036	772063	95.265
9	17.085	11740	2021	0.237
10	17.461	31497	5950	0.634
Total		4964088	808203	100.000