

Sample Name :Bradykinin (1-3)
 Cat No.: RP11130
 Lot No.: E4791519I
 Time Processed :15:31:55
 Month-Day-Year Processed :09/30/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm


<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
33.00	Pumps	Solvent B Conc.	5
33.01	Controller	Stop	

<<Column Performance>>

<Detector A>
 Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment:ZJ19010324

<Chromatogram>



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	5.308	137992	9792	0.732
2	5.587	18329226	834736	97.254
3	6.538	98029	9482	0.520
4	6.838	106335	3841	0.564
5	7.733	11793	561	0.063
6	9.224	7492	599	0.040
7	10.114	19125	428	0.101
8	11.070	3117	248	0.017
9	11.410	2114	218	0.011
10	12.042	32783	1288	0.174
11	12.624	2196	157	0.012
12	13.191	13884	877	0.074
13	13.648	2927	452	0.016
14	13.758	8871	491	0.047
15	14.506	1998	183	0.011

Peak#	Ret. Time	Area	Height	Area%
16	15.358	3646	296	0.019
17	15.880	2743	234	0.015
18	16.307	4621	334	0.025
19	16.730	2366	160	0.013
20	17.244	4779	354	0.025
21	17.707	2753	214	0.015
22	18.166	3505	204	0.019
23	18.711	1941	148	0.010
24	19.169	2962	210	0.016
25	19.589	1170	105	0.006
26	21.544	1622	55	0.009
27	22.221	4677	374	0.025
28	22.633	16383	1392	0.087
29	23.297	4263	329	0.023
30	24.109	11443	587	0.061
Total		18846757	868352	100.000