


Mass Spectrum



Sample Information

User : Gary
Month-Day Processed : 07/03/19
Time Processed : 08:58:17 PM
Injection Volume : 0.3
Sample Name : Parasin I
Cat No. : RP11233
Lot No. : E3316319G
Theoretical MW : 2000.30
Observed MW : 2000.5

Equipment : GK11010007
Interface Bias : +4.5kV
Drying Gas Flow : 5 L/min
T.Flow : 0.2ml/min
B.conc : 50%H2O/50%MeOH

Interface : ESI
Nebulizing Gas Flow : 1.5L/min
CDL Temp : 250
Block Temp : 200